

eKare - Calibration Accuracy Frame

- * Use this frame to verify calibration accuracy of the eKare inSight™ device
- * Upper & Lower Confidence Intervals of 5% define the acceptance range
- * Measure the round, blue circle on this card
- * inSight™ device is accurately calibrated if Length, Width, **AND** Area measurements lie within the intervals identified in the table below

	True (cm)	Acceptance Range (± 5% of True, cm)	
		LCL	UCL
Area	19.635	18.653	20.617
Length	5.00	4.75	5.25
Width	5.00	4.75	5.25

